



CONVERSION TABLE FOR
MODEL 919/3.5 MOISTURE METER

TABLEAU DE CONVERSION POUR
HUMIDIMÈTRE DE MODÈLE 919/3,5

EASTERN CORN

SAMPLE / ÉCHANTILLON

MAÏS DE L'EST

Moisture range to 19.9%

250 g

Marge d'humidité jusqu'à 19,9 %

Meter Reading	TEMPERATURE °C TEMPÉRATURE																				Relevé d'humidité
	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	
	MOISTURE % TENEUR EN EAU																				
52.5	16.6	16.6	16.5	16.4	16.3	16.2	16.1	16.0	16.0	15.9	15.8	15.7	15.6	15.5	15.4	15.4	15.3	15.2	15.1	15.0	52.5
53.0	16.8	16.7	16.6	16.5	16.4	16.3	16.2	16.1	16.1	16.0	15.9	15.8	15.7	15.6	15.5	15.5	15.4	15.3	15.2	15.1	53.0
53.5	16.9	16.8	16.7	16.6	16.5	16.4	16.3	16.2	16.2	16.1	16.0	15.9	15.8	15.7	15.6	15.6	15.5	15.4	15.3	15.2	53.5
54.0	17.0	16.9	16.8	16.7	16.6	16.5	16.4	16.3	16.3	16.2	16.1	16.0	15.9	15.8	15.7	15.7	15.6	15.5	15.4	15.3	54.0
54.5	17.1	17.0	16.9	16.8	16.7	16.6	16.5	16.4	16.4	16.3	16.2	16.1	16.0	15.9	15.8	15.8	15.7	15.6	15.5	15.4	54.5
55.0	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.5	16.5	16.4	16.3	16.2	16.1	16.0	15.9	15.9	15.8	15.7	15.6	15.5	55.0
55.5	17.3	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.6	16.5	16.4	16.3	16.2	16.1	16.0	16.0	15.9	15.8	15.7	15.6	55.5
56.0	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.8	16.7	16.6	16.5	16.4	16.3	16.2	16.1	16.1	16.0	15.9	15.8	15.7	56.0
56.5	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.5	16.4	16.3	16.2	16.1	16.2	16.1	16.0	15.9	15.8	56.5
57.0	17.6	17.5	17.4	17.3	17.2	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.5	16.4	16.3	16.3	16.2	16.1	16.0	15.9	57.0
57.5	17.7	17.6	17.5	17.4	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.5	16.5	16.4	16.3	16.2	16.1	16.0	57.5
58.0	17.8	17.7	17.6	17.6	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.6	16.5	16.4	16.3	16.2	16.1	58.0
58.5	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.7	16.7	16.6	16.5	16.4	16.3	16.2	58.5
59.0	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.9	16.8	16.7	16.6	16.5	16.4	16.3	59.0
59.5	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.0	17.0	16.9	16.8	16.7	16.6	16.5	16.4	59.5
60.0	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.2	17.1	17.0	16.9	16.8	16.7	16.6	16.5	60.0
60.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.7	16.6	60.5
61.0	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.8	16.7	61.0
61.5	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.0	16.9	16.8	61.5
62.0	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.0	16.9	62.0
62.5	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.2	17.1	17.1	62.5
63.0	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	17.4	17.3	17.3	17.2	63.0
63.5	19.2	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.6	17.5	17.4	17.3	63.5
64.0	19.3	19.2	19.1	19.0	18.9	18.8	18.6	18.5	18.4	18.3	18.2	18.1	18.0	18.0	17.9	17.8	17.7	17.6	17.5	17.4	64.0
64.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	17.5	64.5
65.0	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	17.6	65.0
65.5	19.7	19.6	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	17.7	65.5
66.0	19.8	19.7	19.6	19.5	19.4	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	17.8	66.0
66.5	19.9	19.8	19.7	19.6	19.5	19.4	19.3	19.2	19.1	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	18.0	17.9	66.5
67.0		19.9	19.8	19.7	19.6	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	67.0
67.5			19.9	19.7	19.6	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	18.1	67.5
68.0				19.9	19.8	19.6	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	18.2	68.0
68.5					19.9	19.8	19.7	19.6	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	18.3	68.5
69.0						19.9	19.8	19.7	19.6	19.5	19.4	19.3	19.1	19.0	18.9	18.8	18.7	18.6	18.5	18.4	69.0
69.5							19.9	19.8	19.7	19.6	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.7	18.6	18.5	69.5
70.0									19.8	19.7	19.6	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	18.7	70.0
70.5										19.9	19.7	19.6	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	70.5
71.0											19.9	19.7	19.5	19.4	19.3	19.2	19.1	19.0	18.9	18.8	71.0
71.5												19.9	19.8	19.7	19.6	19.5	19.3	19.2	19.1	19.0	71.5
72.0													19.9	19.8	19.7	19.6	19.5	19.4	19.3	19.2	72.0
72.5														19.9	19.8	19.7	19.6	19.5	19.4	19.3	72.5
73.0															19.8	19.6	19.5	19.4	19.3	19.2	73.0
73.5																19.9	19.8	19.6	19.5	19.4	73.5
74.0																	19.9	19.8	19.7	19.6	74.0
74.5																		19.9	19.8	19.7	74.5

TABLE NO. 6 - JULY 2000
TABLEAU N° 6 - JUILLET 2000

USE CLEANED SAMPLE
UTILISER UN ÉCHANTILLON NETTOYÉ

CALIBRATION: 103°C, 72 h, Air Oven
ÉTALONNAGE : 103 °C, 72 h, Étuve